

Dell Medical Certification Statement
01/22/05

Dell products are designed to meet product certification requirements for use in general administrative and information processing environments within medical facilities.

Dell products are not designed or certified for direct patient contact or for use as medical equipment or medical devices. In addition, Dell products are not designed or certified for use within 1.83 meters (6 feet) of a patient or in explosive environments where flammable gases such as anesthesia or oxygen may be present.

The table below provides general compliance information regarding Dell products¹. Please check the actual product specifications for compliance information on specific products. You should consult with your safety team to determine the applicability of these and any other standards based on your intended use of any product.

Dell Product	NFPA 99	UL 60601-1, UL 2601-1, IEC 60601-1
Most Currently shipping Dell OptiPlex, GX and SX Family Desktops; Dell Latitude Notebooks and Dell Precision Workstations	Comply with the “less than 300 micro-amp” chassis leakage requirements of NFPA 99:2002 section 8.4.1.3.5.1	Not certified.
Dell Dimension and Inspiron products. Dell Monitors and Flat Panel Displays	Not compliant. ²	Not certified.
Dell Printers	Not compliant. ²	Not certified.
Dell Peripherals (Axim etc.)	Not compliant. ²	Not certified.

¹ Certain third-party manufacturers may include Dell products as part of a “mobile cart” or other solution. Dell does not verify or evaluate any medical certification statements made by such manufacturers.

² Customer may consider using an appropriate isolation transformer or a medical grade Un-interruptible Power Supply (UPS) in environments that require less than 300 micro-amp chassis leakage